



IMPORTANT CLINICAL LABORATORY MEMORANDUM

Date: 3/17/2026

To: CAMC Providers
CAMC Staff
CAMC LabWorks Clients

From: CAMC Vandalia Health South Region Laboratories

Milton Plata, MD
Darlene Gruetter, MD

TOPIC: Discrepant Total T4 (Thyroxine) results

EFFECTIVE: 3/17/2026

Beckman Coulter recently released a manufacturer Urgent Medical Device Recall that states Total T4 (Thyroxine) results could potentially be falsely decreased by up to 23% on Unicel Dxl 600 and UnicelDxl 800 instruments for specific lots of calibrators distributed between January 2026 through March 2026. If decreased Total T4 results are suspected, please assess Total T4 results in conjunction with other patient values and clinical presentation.

Beckman Coulter is no longer distributing affected lot numbers of Total T4 calibrators.

This notice applies to testing performed at the following CAMC hospital facilities: Memorial and Teays Valley.

For any further questions please contact the performing laboratory.

Memorial Lab: Megan Peaytt 304-388-4184

Teays Valley Lab: Angela Warner 304-757-1769